

1.00361.0001

Microscopy

ISOSLIDE® Reticulin Control Slides

with reference tissue for the detection of reticular fibers in histology

For professional use only



In Vitro Diagnostic Medical Device



Intended purpose

These "ISOSLIDE® Reticulin Control Slides - with reference tissue for the detection of reticular fibers in histology" are used for human-medical cell diagnosis and serve the purpose of the histological investigation and quality control of sample material of human origin. They are ready-to-use glass slides with paraffin sections that contain the typical structures of the corresponding test method. They, when used together with other in vitro diagnostic products from our portfolio, make target structures evaluable for diagnostic purposes (by fixing, embedding, staining, counterstaining, mounting) in human-histological specimen material, for example histological sections of e.g. the kidney or the liver.

The ISOSLIDE® Reticulin Control Slides are prepared using sections of suitable animal and animal material, that clearly visualize reticular fibers. Each package of ISOSLIDE® control slides contains 25 slides. The tissue is applied onto the slide next to the white printed labelling field. One slide is already stained with the reference method and is used for comparison. The 24 unstained slides are stained according to the corresponding protocol or with the laboratory's own protocol.

After the staining of the control slides, the staining result is to be compared with the laboratory material and the pre-stained control slide.

Due to variable procedures in sampling, fixation, histoprocessing, paraffinization and equipment, no statements about these processing steps can be made from the ISOSLIDE® control slides. The ISOSLIDE® control slides are not quality-controlled for their application with molecular pathology methods and IHC.

Unstained structures are relatively low in contrast and are extremely difficult to distinguish under the light microscope. The images created using the staining solutions help the authorized and qualified investigator to better define the form and structure in such cases. Further examinations may be necessary to reach a definitive diagnosis.

Principle

In the reticulin silver plating method acc. to Gordon & Sweets silver salts are deposited on the reticulin fibers, which then become visible as black structures. In this method the material is oxidized with potassium permanganate and reduced with formaldehyde. Gold chloride serves to better stabilize the metal compound, which in turn makes the results more intense.

Sample material

Paraffin sections with approx. 3 - 4 µm, from tissue samples of animal origin. The tissue was fixed with neutral buffered formalin.

Reagents

Cat. No. 1.00361.0001
ISOSLIDE® Reticulin Control Slides
with reference tissue for the detection of reticular fibers in histology

Package components:

24 unstained control slides with reticular fibers
1 stained control slide (showing only reticulin) - stained with Reticulin silver plating kit acc. to Gordon & Sweets

Also required:

Cat. No. 100121	Nuclear fast red-aluminum sulfate solution 0.1 % for microscopy	500 ml
Cat. No. 100251	Reticulin silver plating kit acc. to Gordon & Sweets	1 set

Sample preparation

The sampling must be performed by qualified personnel.

All samples must be treated using state-of-the-art technology.

All samples must be clearly labeled.

Suitable instruments must be used for taking samples and their preparation. Follow the manufacturer's instructions for application / use.

When using the corresponding auxiliary reagents, the corresponding instructions for use must be observed.

Deparaffinize sections in the conventional manner and rehydrate in a descending alcohol series.

Procedure

Analogous to Cat. No. 100251 - Reticulin silver plating kit acc. to Gordon & Sweets:

The staining kit includes 8 reagents contained in 30-ml dropper bottles. The reagents are dropped onto the slides according to the staining protocol. The staining itself is performed on a staining rack made of plexiglass or coated with plastic.

It is also possible to stain slides with nuclear fast red solution to demonstrate the nuclei and other structures; the procedure is described in the instructions for use of the Reticulin silver plating kit acc. to Gordon & Sweets.

It is important to rinse the slides intensively with distilled water.

Do not use metal tweezers and do not allow any other metal objects to come into contact with the slides.

After dehydration (ascending alcohol series) and clarification with xylene or Neo-Clear®, histological slides can be covered with non-aqueous mounting agents (e.g. Entellan® new, Neo-Mount®) and a cover glass and can then be stored.

Result

Reticular fibers black

With nuclear fast red solution

Nuclei	red
Background	red
Collagen	red

Trouble-shooting

Silver-staining techniques can be difficult and require special care during the procedure.

Microscopic image	Possible cause	Remedy
Poor visualization of reticulin fibers	Gold chloride solution incubated too long	See "Procedure"
Poor visualization of reticulin fibers	Slide insufficiently covered with reagent	Increase the number of drops or use a histomarker
Microscopic image	Possible cause	Remedy
Poor or no visualization of reticulin fibers	Incorrect storage / shelf-life of the kit	Observe storage and shelf-life
Unclean slide / precipitation	Insufficiently rinsed	See "Procedure"
Reticulin fibers stained too intensely	Gold chloride solution not incubated long enough	See "Procedure"

Technical notes

The microscope used should meet the requirements of a medical diagnostic laboratory.

When using histoprocessor systems or automatic staining systems, please follow the instructions for use supplied by the supplier of the system and software.

Diagnostics

Diagnoses are to be made only by authorized and qualified personnel. Valid nomenclatures must be used.

This method can be supplementarily used in human diagnostics.

Further tests must be selected and implemented according to recognized methods.

Suitable controls should be conducted with each application in order to avoid an incorrect result.

Storage

Store the ISOSLIDE® Reticulin Control Slides - with reference tissue for the detection of reticular fibers in histology at +15 °C to +25 °C.

Shelf-life

The ISOSLIDE® Reticulin Control Slides - with reference tissue for the detection of reticular fibers in histology can be used until the stated expiry date.

The package can be used up to the stated expiry date when stored at +15 °C to +25 °C.

Capacity

The package is sufficient for 24 applications.

Additional instructions

For professional use only.

In order to avoid errors, the application must be carried out by qualified personnel only.

National guidelines for work safety and quality assurance must be followed. Microscopes equipped according to the standard must be used.

Protection against infection

Effective measures must be taken to protect against infection in line with laboratory guidelines.

Instructions for disposal

The kit and any unused control slides must be disposed of in accordance with the current disposal guidelines.

Used solutions and solutions that are past their shelf-life must be disposed of as special waste in accordance with local guidelines. Information on disposal can be obtained under the Quick Link "Hints for Disposal of Microscopy Products" at www.microscopy-products.com. Within the EU the currently applicable REGULATION (EC) No 1272/2008 on classification, labelling and packaging of substances and mixtures, amending and repealing Directives 67/548/EEC and 1999/45/EC, and amending Regulation (EC) No 1907/2006 applies.

Auxiliary reagents

Cat. No. 103699	Immersion oil Type N acc. to ISO 8036 for microscopy	100-ml dropping bottle
Cat. No. 104699	Immersion oil for microscopy	100-ml dropping bottle, 100 ml, 500 ml
Cat. No. 107961	Entellan® new rapid mounting medium for microscopy	100 ml, 500 ml, 1 l
Cat. No. 108298	Xylene (isomeric mixture) for histology	4 l
Cat. No. 109016	Neo-Mount® anhydrous mounting medium for microscopy	100-ml dropping bottle, 500 ml
Cat. No. 109843	Neo-Clear® (xylene substitute) for microscopy	5 l

Hazard classification

Cat. No. 1.00361.0001

Please observe the hazard classification printed on the label and the information given in the safety data sheet.

The safety data sheet is available on the website and on request.

Main product components

Cat. No. 1.00361.0001

Paraffin sections with animal tissue: 25 slides

Other IVD products

Cat. No. 100380	ISOSLIDE® Iron Control Slides with reference tissue for the detection of free iron in histological tissue	25 tests
Cat. No. 100408	ISOSLIDE® PAS Control Slides with reference tissue for the detection of polysaccharides in histological tissue	25 tests
Cat. No. 100425	ISOSLIDE® Alcian blue, pH 2.5 Control Slides with reference tissue for the detection of acid mucosubstances in histological tissue	25 tests
Cat. No. 102472	ISOSLIDE® Warthin-Starry Control Slides with reference tissue for the detection of Helicobacter pylori and Spirochetes in histological tissue	25 tests
Cat. No. 102473	ISOSLIDE® Methenamine Control Slides with reference tissue for the detection of argent-affine structures in histological tissue	25 tests
Cat. No. 102560	ISOSLIDE® AFB Control Slides with reference tissue for the detection of acid-fast bacteria in histological tissue	25 tests

General remark

If during the use of this device or as a result of its use, a serious incident has occurred, please report it to the manufacturer and/or its authorised representative and to your national authority.

Literature

1. Romeis - Mikroskopische Technik, Editors: Maria Mulisch, Ulrich Welsch, 2015, Springer Spektrum, 19. Auflage
2. Welsch Sobotta - Lehrbuch Histologie, Editor: Ulrich Welsch, 2006, ELSEVIER Urban&Fischer, 2. Auflage
3. Histotechnik, Gudrun Lang, 2013 Springer Verlag, 2. Auflage
4. Theory and Practice of Histological Techniques, John D Bancroft, Marilyn Gamble, 2008, Churchill Livingstone ELSEVIER, 6th Edition
5. Laboratory Manual of Histochemistry, Linda L. Vacca, 1985, Raven Press
6. Basiswissen Histologie und Zytologie, Karl Heinz Stein, Hellmut Flenker, 1998, uZv, 2. Auflage
7. Histological & Histochemical Methods: Theory & Practice, J. A. Kiernan, 1990, Pergamon Press, 2nd Edition
8. Histological and Histochemical Methods, Theory and practice, J. A. Kiernan, 2015, Scion Publishing Ltd, 5th Edition



Consult instructions for use



Manufacturer



Catalog number



Batch code



Caution, consult accompanying documents



Use by YYYY-MM-DD



Temperature limitation

Status: 2021-Jan-19

Merck KGaA, 64271 Darmstadt, Germany,
Tel. +49(0)6151 72-2440
www.microscopy-products.com

EMD Millipore Corporation, 400 Summit Drive
Burlington MA 01803, USA, Tel. +1-978-715-4321
Sigma-Aldrich Canada Co. or Millipore (Canada) Ltd.
2149 Winston Park, Dr. Oakville, Ontario, L6H 6J8
Phone: +1 800-565-1400

